


Nude Nicotine Analytics

8677 Villa La Jolla #1236
La Jolla, CA 92037
(858) 216-2044
www.nudenicotine.com
admin@nudenicotine.com



Nude Nicotine – GC/MS Analysis Results

| Customer Name: | Date of Sample Receipt: | External Lot Number: |
|----------------|---------------------------------|-------------------------|
| You Got eJuice | February 9 th , 2016 | N/A |
| Product Name: | Date of Analysis: | NN Analytics Sample ID: |
| Yogurt | February 9 th , 2016 | YG14 |

Scope of Method
to be Performed:

Diketone Trio Analysis:

Aliquot is diluted 10:1 with methanol containing a known amount of butyric-d7 acid as an internal standard. The aliquot is then capped, mixed, and injected into the GC/MS system. Analyte identification is confirmed by no less than 3 unique ion fragments by the mass selective detector. Results are quantitated utilizing a calibration curve generated from certified reference materials.

| Analytical Compound | Description | CAS # | Limit of Detection (LOD) | Results |
|-------------------------------------|-------------|----------|--------------------------|------------------------|
| 2,3-Butanedione (diacetyl) | Diketone | 431-03-8 | 1.08ug/mL | ND |
| 2,3-pentanedione (acetyl propionyl) | Diketone | 600-14-6 | 1.26ug/mL | <LOQ |
| 3-Hydroxy-2-butanone (acetoin) | Ketone | 513-86-0 | 1.01ug/mL | 58.8ug/mL ± 0.059ug/mL |

ND = Not Detected / 1ug/mL = 1ppm / 1000ug/mL = 1mg/mL

The above results are guaranteed accurate and precise to the best of our ability. Data has been checked and in accordance with Nude Nicotine's approved analytical protocols. I certify that all results in this report are adherent to this statement:

| GC/MS Operator and Analytical Report Prepared by: | Signature & Date: |
|---|-------------------|
| J. Rubenstein | |

Notes: Samples were received in good condition with factory seals intact. Chain of custody was properly maintained during the entirety of the sampling and receipt process.

This document was drafted as a .pdf and only contains a single page.

